

LASER SPECTROSCOPY

Using active FTIR to distinguish between multiple gases

Christopher G. Leburn, Oguzhan Kara, and Derryck T. Reid

An active long-range FTIR spectroscopy system can acquire high-resolution gas absorption spectra such as hydrocarbon emissions.

Fugitive hydrocarbon emissions are estimated to cost the energy sector \$5 billion per year and account for 12% of greenhouse gas emissions—they are also thought to jeopardize safety and public health, as well as a key factor in climate change. Whether it's to enable industrial sites close to communities to conduct 24-hour continuous monitoring, or to assess efficiencies in combustion engines, markets including oil and gas, landfill, and agriculture are driving a growing requirement for high-resolution gas-detection solutions that are compact, portable, and affordable.

Analysing different gases can be achieved in a number of different ways. Differential absorption lidar (DIAL) is considered to be one of the most advanced techniques. With a range of 500 m, this technique directs high-energy laser light into the atmosphere, which is returned to a ground-based detector by weak scattering from airborne particles. Unfortunately, DIAL systems are complex, costly to run, with a large footprint and some systems are often housed on a 18 wheeler truck.

By contrast, Fourier-transform infrared (FTIR) spectroscopy is naturally broadband and offers far wider coverage than DIAL.

Fugitive hydrocarbon emissions are estimated to cost the energy sector \$5 billion per year and account for 12% of greenhouse gas emissions.

In an open-path setup, FTIR has the capability to detect hundreds of atmospheric gases and, with a small system footprint comparable to a briefcase, is much more portable than a DIAL system. Open-path FTIR typically uses thermal sources to quantify hydrocarbon emissions, but with typical commercial solutions having resolutions of 0.5 cm^{-1} , it is difficult to separate out multiple species when they are spectrally overlapped. In addition, in-the-field FTIR spectroscopy using thermal sources typically requires high-quality retroreflecting targets to direct the source light back towards the detector.

Laser-based active FTIR spectroscopy offers higher resolution, providing the capability to distinguish similar gases such as methane and ethane.




FIGURE 1. These idler spectra are produced by fan-out-grating tuning of the Chromacity OPO. The spectral shapes are determined by the phase-matching characteristics of the OPO crystal and by the OPO pump-laser spectrum. Water absorption lines are visible at shorter wavelengths.

LASER SPECTROSCOPY

Using active FTIR to distinguish between multiple gases

To achieve this in the atmosphere, a broadband light source is required with a brightness high enough to be effective over long distances—such a light source was not previously available on the market. Quantum-cascade laser (QCL) technologies have been used for stand-off detection measurements of water vapor, methane, nitrous oxide, and hydrogen peroxide. However, whilst a cheaper alternative than DIAL, QCLs provide only a narrow linewidth, which limits their capability when it comes to the detection of multiple species.

Researchers at Heriot-Watt University have been working with ultrafast laser specialist Chromacity (both in Edinburgh, Scotland) to develop an eye-safe active FTIR spectroscopy system capable of acquiring sub-0.1-cm⁻¹ resolution gas absorption spectra from simple targets at ranges exceeding 70 m. This is a first step towards a real-time solution for simultaneously quantifying multiple gases over several hundred meters.





FIGURE 2. The OPO, telescope, and scanning Michelson interferometer are seen on a 60 x 90 cm breadboard (a); the layout of the Fourier-transform spectrometer is seen in (b).



THE GAS-ANALYSING SYSTEM

The experiment is broken into three main elements: the broadband light source, the spectrometer (consisting of interferometer and detection system), and a computer algorithm to extract the data.

The light source was the Chromacity broadband, ultrafast optical parametric oscillator (OPO). This particular model was optimized to provide tunability from 2800 to 3900 nm, so that

the strongest absorptions in methane and ethane (3.1–3.5 μm) could be detected and identified (see Fig. 1). This high-brightness source delivered average powers greater than 300 mW as a 1cm diameter beam was passed into the spectrometer.

Light from the Chromacity OPO was first coupled into a scanning Michelson interferometer before being launched into free space (see Fig. 2). The scattered signal returned from the target was subsequently collected by a 6-in. f/4 Newtonian telescope and detected using an indium antimonide (InSb) liquid-nitrogen-cooled photodiode. Light from the OPO was launched along an optical axis co-aligned with the telescope's field of view using a small 45° steering mirror situated directly before the secondary mirror of the telescope. The scanning interferometer operated at 1 Hz and achieved a typical resolution of 0.05 cm⁻¹, which is sufficient to resolve the narrow and complex absorption-line structure of light molecules like water, methane and ethane.




FIGURE 3. An example spectrum (a) is shown from a set of 45 collected at 30 m range from a rough aluminum-foil target. Measured concentrations at 10, 20, 25, and 30 m are plotted (b), with data points showing the average values from approximately 45 spectra each, and the error bars showing the ± 1 standard deviation range. Concentrations were measured to be 1.15% water, 1.37% ethane, and 1860 ppb methane.

To deconvolve the information received from the spectrometer, a patent-pending computer algorithm was developed. Quantitative open-path spectroscopy requires either a reliable reference spectrum or a method of inferring the original illumination spectrum, and this problem has been resolved in different ways. The approach used here resulted in the retrieval of an illumination spectrum, which represents the OPO output spectrum prior to undergoing atmospheric absorption.

MULTI-SPECIES MEASUREMENTS AT 30 M RANGE

The first series of experiments was designed to establish the ability to measure multiple spectrally overlapping species simultaneously. Indoor measurements were performed using a rough aluminum-foil target placed at a range of up to 30 m from the FTIR system. Light emitted from the OPO entered a 20 cm long gas cell containing a 1.5±0.15% ethane-in-air mixture. This cell was situated directly in front of the secondary mirror of the telescope.

Figure 3 gives an example of a single measured spectrum (no averaging) exhibiting densely packed absorption lines from water, methane, and ethane, as well as continuum absorption from ethane, which suppresses the overall spectral intensity.

The algorithm retrieves the effective illumination spectrum (see Fig. 3a, dashed line), which represents the OPO output spectrum prior to undergoing atmospheric absorption.

Laser-based active FTIR spectroscopy offers higher resolution, providing the capability to distinguish between multiple gases.

REAL-TIME METHANE EMISSION MEASUREMENT AT 70 M RANGE

The next phase of the project extended the range of detection and also performed a real-time measurement of a simulated gas leak. A 78 m corridor was used at the test site. A 2% methane:air mix was released for 100 s at a rate of 103 μg s⁻¹ at a distance of 65 m from the Chromacity OPO. During this experiment, the ethane cell was not present. Spectra were recorded every seven seconds and repeated in the same way as the previous experiment to provide concentrations of water, methane, and ethane.

Figure 4a shows an example of a spectrum recorded without averaging at 70 m range and at a moment close to the peak methane emission. The inset of Figure 4a shows the correspondence to the best-fit PNNL database across the 3.18–3.21 μm region. In contrast to ethane, methane and water show very little continuum absorption




FIGURE 4. This example spectrum (a) is from a set of 51 collected at 70 m range from a rough aluminum-foil target. In the example, the best-fit concentrations were 0.77% water, 12650 ppb methane, and 0% ethane. Concentrations of methane (blue) and water (red) were measured before, during, and after the methane release (b), showing a recovery to near-background levels after 150 s.

LASER SPECTROSCOPY

Using active FTIR to distinguish between multiple gases

under these experimental conditions (20°C, 101,800 Pa), so the inferred illumination spectrum closely follows the envelope of the measured spectrum. Figure 4b presents the measured water and methane concentrations over 400 s, showing the methane concentration rising from background levels (around 1900 ppb) to a peak of around 13,000 ppb before returning to near the original value as the gas disperses. Prior to the methane release, the root mean squared (RMS) variation of the measured concentration of background methane at this range was <100 ppb. Water showed more variability, which is assumed to be associated with convection effects and environmental variations.

This experimental set-up offers several advantages over previously reported open-path analysing of atmospheric species. The measurements produced here were carried out with very simple, alignment-free targets, typically a coarsely positioned sheet of aluminum foil at a range of up to 70 m, but also with returns from >30 m available from paper, laminate, or similar targets.

EMISSION MEASUREMENT BEYOND 70 M

During these experiments, the distance of the measurements have been constrained by the dimensions of the available free space. Based on the high brightness of the Chromacity OPO, the signal-to-noise performance infers that measurements beyond 70 m will be possible with no modification to the current setup.

The ability to use a simple topographic target increases the practicality of the system. Since there is no need to maintain the beam on a small retroreflector, the beam-pointing accuracy needed is reduced and, in some cases, eliminates the need for a remote target. In more recent outdoor experiments, the measurements were extended at a 3.5 μm wavelength to more than 200 m, using only the aluminum tailgate of a truck as the retroreflecting target.

A highlight of this methodology is the ability to extract concentration data from a single spectrum with no need for averaging, which provides a real-time and quantitative monitoring capability making it suitable for detecting even weak, localized emissions.

The resolution of 0.05 cm^{-1} shows that it is possible to extract concentration measurements from complex, spectrally overlapping multiple species. Combined with the real-time monitoring capabilities, this feature could permit the observation of correlations between different hydrocarbon gas concentrations, providing insights into their origins. For example, separate petrochemical methane contributions (accompanied by weak ethane signatures) can be separated from biogenic sources (such as cattle, landfill, etc.), which produce only methane.

The entire combined system, including the Chromacity OPO and spectrometer, fits on a 60 × 90 cm breadboard, making it more portable and thus easy to transport. The system also has the ability to quantify many other chemical signatures. These features open up significant opportunities for the development of a cost-effective and highly portable environmental multi-gas analysing solution.



FIGURE 5. The Chromacity OPO

ABOUT THE AUTHORS

Christopher G. Leburn is at Chromacity
Oguzhan Kara and Derryck T. Reid
were at Heriot-Watt University.

All located in Edinburgh, Scotland.

e-mail c.g.leburn@chromacitylasers.com
web www.chromacitylasers.com